

## KAMGO LAB WATER

## Certificate Of Analysis

## LC-MS GRADE WATER


Batch No:202306-00100-LCMS

Mfg.Date:06/2023

Parameter	Results
Aluminium	<0.5 ppm
Calcium	<0.1 ppm
Iron	<0.1 ppm
Magnesium	<0.1 ppm
Nickel	<0.02 ppm
Potassium	<0.1 ppm
Sodium	<0.1 ppm
Silver (Ag)	<1 ppm
Barium (Ba)	<0.1 ppm
Cadmium (Cd)	<0.05 ppm
Cobalt (Co)	<0.02 ppm
Chromium (Cr)	<0.02 ppm
Copper (Cu)	<0.02 ppm
Manganese (Mn)	<0.02 ppm
Lead (Pb)	<0.1 ppm
Tin (Sn)	<0.1 ppm
Zinc (Zn)	<0.1 ppm
Sulfate (SO <sub>4</sub> )	<0.1 ppm
Fluoride(F)	<0.01 ppm
Nitrate(NO <sub>3</sub> )	<0.1 ppm
HPLC Gradient at 210 nm	<5 mAU
HPLC Gradient at 254 nm	<1 mAU
Suitability for the LC-MS	COMPLYING
Absorbance at 210 nm	<0.022
Absorbance at 230 nm	<0.004
Transmittance at 210nm	>95%
Transmittance at 230nm	>99%
Fluorescence (chinin) at 254nm	<1 ppb
Fluorescence (chinin) at 365nm	<1ppb

Filtered through 0.2µm pore size filter

KAMGO TECHNOLOGIES INDIA PVT LTD



KAMGO LAB WATER

Certificate Of Analysis

Kamgo Technologies India Pvt Ltd  
Authorized Signature



Note: Results arrived based on testing at NABL Laboratory