



KAMGO TECHNOLOGIES INDIA PVT LTD[®]
KAMGO LAB WATER
Certificate Of Analysis

LC-MS GRADE WATER

Batch No: 2025004-LCMS

Mfg. Date: 08/2025

Exp Date: 07/2026

Parameter	Results
Aluminium	<0.5 ppm
Calcium	<0.1 ppm
Iron	<0.1 ppm
Magnesium	<0.1 ppm
Nickel	<0.02 ppm
Potassium	<0.1 ppm
Sodium	<0.1 ppm
Silver (Ag)	<1 ppm
Barium (Ba)	<0.1 ppm
Cadmium (Cd)	<0.05 ppm
Cobalt (Co)	<0.02 ppm
Chromium (Cr)	<0.02 ppm
Copper (Cu)	<0.02 ppm
Manganese (Mn)	<0.02 ppm
Lead (Pb)	<0.1 ppm
Tin (Sn)	<0.1 ppm
Zinc (Zn)	<0.1 ppm
Sulfate (SO ₄)	<0.1 ppm
Fluoride(F)	<0.01 ppm
Nitrate(NO ₃)	<0.1 ppm
HPLC Gradient at 210 nm	<5 mAU
HPLC Gradient at 254 nm	<1 mAU
Suitability for the LC-MS	COMPLYING
Absorbance at 210 nm	<0.022
Absorbance at 230 nm	<0.004
Transmittance at 210nm	>95%
Transmittance at 230nm	>99%

Filtered through 0.2µm pore size filter

Kamgo Towers, No 141, Nehru Nagar, 9th Link Street, Kottivakkam, Chennai - 600041

www.kamgotech.com /info@kamgotech.com



KAMGO TECHNOLOGIES INDIA PVT LTD[®]
KAMGO LAB WATER

Kamgo Technologies India Pvt Ltd



Authorized Signature

Note: Results arrived based on testing at NABL Laboratory

Kamgo Towers, No 141, Nehru Nagar, 9th Link Street, Kottivakkam, Chennai - 600041

www.kamgotech.com / info@kamgotech.com